


CERTIFICATE OF ANALYSIS

prepared for: REVIVE CBD 5055 S KIPLING PKWY C5 LITTLETON, CO 80127

REVIVE CBD PET TINCTURE 300mg 30ml

Reported:13-Sep-2019Method:TM14Type:ConcentrateTest:Potency	Batch ID:		Test ID:	4247801.0013
	Reported:	13-Sep-2019	Method:	TM14
Test: Potency	Туре:	Concentrate		
	Test:	Potency		

CANNABINOID PROFILE



Compound	LOQ (%)	Result (%)	Result (mg/g)
Delta 9-Tetrahydrocannabinolic acid (THCA-A)	0.01	0.00	0.0
Delta 9-Tetrahydrocannabinol (Delta 9THC)	0.00	0.06	0.6
Cannabidiolic acid (CBDA)	0.01	0.12	1.2
Cannabidiol (CBD)	0.00	1.29	12.9
Delta 8-Tetrahydrocannabinol (Delta 8THC)	0.00	0.00	0.0
Cannabinolic Acid (CBNA)	0.01	0.00	0.0
Cannabinol (CBN)	0.00	0.00	0.0
Cannabigerolic acid (CBGA)	0.01	0.00	0.0
Cannabigerol (CBG)	0.00	0.04	0.4
Tetrahydrocannabivarinic Acid (THCVA)	0.00	0.00	0.0
Tetrahydrocannabivarin (THCV)	0.00	0.00	0.0
Cannabidivarinic Acid (CBDVA)	0.01	0.00	0.0
Cannabidivarin (CBDV)	0.00	0.00	0.0
Cannabichromenic Acid (CBCA)	0.00	0.00	0.0
Cannabichromene (CBC)	0.01	0.05	0.5
Total Cannabinoids	1.56	15.60	
Total Potential THC**	0.06	0.60	
Total Potential CBD**		1.40	13.95

NOTES:

N/A

%=% (w/w) = Percent (Weight of Analyte / Weight of Product)

* Total Cannabinoids result reflects the absolute sum of all cannabinoids detected.
** Total Potential THC/CBD is calculated using the following formulas to take into account the loss of a carboxyl group during decarboxylation step.

Total THC = THC + (THCa *(0.877)) and Total CBD = CBD + (CBDa *(0.877))

FINAL APPROVAL

Daniel Wartanna

Daniel Weidensaul 13-Sep-2019 5:01 PM

PREPARED BY / DATE



Greg Zimpfer 13-Sep-2019 5:19 PM

APPROVED BY / DATE

Testing results are based solely upon the sample submitted to Botanacor Laboratories, LLC, in the condition it was received. Botanacor Laboratories, LLC warrants that all analytical work is conducted professionally in accordance with all applicable standard laboratory practices using validated methods. Data was generated using an unbroken chain of comparison to NIST traceable Reference Standards and Certified Reference Materials. This report may not be reproduced, except in full, without the written approval of Botanacor Laboratories, LLC. ISO/IEC 17025:2005 Accredited A2LA Certificate Number 4329.02



Botanacor Laboratories™, All Rights Reserved | 1001 S. Galapago St., Denver, CO 80223 | 888.800.8223 | www.Botanacor.com